



# IEH ANALYTICAL LABORATORIES

LABORATORY & CONSULTING SERVICES

3927 AURORA AVENUE NORTH, SEATTLE, WA 98103

PHONE: (206) 632-2715 FAX: (206) 632-2417

<b>CASE FILE NUMBER:</b>	1765473	<b>PAGE 1</b>
<b>REPORT DATE:</b>	03/16/26	
<b>DATE SAMPLED:</b>	02/23/26	<b>DATE RECEIVED:</b> 02/24/26
<b>FINAL REPORT, LABORATORY ANALYSIS OF SELECTED PARAMETERS ON WATER SAMPLES FROM MLIRD</b>		

## CASE NARRATIVE

Four water samples were received by the laboratory in good condition and analyzed according to the chain of custody. No difficulties were encountered in the preparation or analysis of these samples. Sample data follows while QA/QC data is contained on the subsequent page.

## SAMPLE DATA

SAMPLE ID	TOTAL-P (mg/L)
TS-2 (Grab Sample)	0.067
TS-5a (0.5m Below Surface)	0.028
TS-6a (0.5m Below Surface)	0.026
TS-14 (Grab Sample)	0.194



# IEH ANALYTICAL LABORATORIES

LABORATORY & CONSULTING SERVICES

3927 AURORA AVENUE NORTH, SEATTLE, WA 98103

PHONE: (206) 632-2715 FAX: (206) 632-2417

<b>CASE FILE NUMBER:</b>	1765473	<b>PAGE 2</b>
<b>REPORT DATE:</b>	03/16/26	
<b>DATE SAMPLED:</b>	02/23/26	<b>DATE RECEIVED:</b> 02/24/26
<b>FINAL REPORT, LABORATORY ANALYSIS OF SELECTED PARAMETERS ON WATER SAMPLES FROM MLIRD</b>		

## QA/QC DATA

QC PARAMETER	TOTAL-P (mg/L)
METHOD	EPA 365.1
DATE ANALYZED	03/03/26
REPORTING LIMIT	0.002
DUPLICATE	
SAMPLE ID	BATCH
ORIGINAL	0.012
DUPLICATE	0.011
RPD	8.14%
SPIKE SAMPLE	
SAMPLE ID	BATCH
ORIGINAL	0.012
SPIKED SAMPLE	0.065
SPIKE ADDED	0.050
% RECOVERY	107.40%
QC CHECK	
FOUND	0.094
TRUE	0.094
% RECOVERY	100.00%
BLANK	<0.002

RPD = RELATIVE PERCENT DIFFERENCE.  
NA = NOT APPLICABLE OR NOT AVAILABLE.  
NC = NOT CALCULABLE DUE TO ONE OR MORE VALUES BEING BELOW THE DETECTION LIMIT.  
OR = RECOVERY NOT CALCULABLE DUE TO SPIKE SAMPLE OUT OF RANGE OR SPIKE TOO LOW RELATIVE TO SAMPLE CONCENTRATION.

SUBMITTED BY:

Damien Gadomski  
Project Manager